



# Fever InSight® Thermometer

Model V916C-AUSV1

Probe Tip

Battery Compartment (on back side)

On/Off Button

Large Display with Fever InSight® backlight

This thermometer should be used by adults only. Do not leave the device unattended with children under the age of 12.

\* Uses CR1632 or equivalent battery

**Follow Instructions for Use. This document provides important product operation and safety information regarding this thermometer. Please read this document thoroughly before using the device and keep for future reference.**

Please read these instructions carefully to ensure accurate temperature measurements and safe operation. When used properly, this product will give you years of accurate, dependable service.

## IMPORTANT SAFETY INFORMATION

- High, prolonged fever requires medical attention especially for young children. Please contact your physician.
- Do not allow children to take their temperatures unattended.
- Do not allow children to walk or run during temperature taking.
- The thermometer is only suitable for the measurement of body temperature. The use of this thermometer is not intended as a substitute for consultation with your doctor.
- The thermometer is initially calibrated at the time of manufacture. If this thermometer is used according to the use instructions, the accuracy of measurements will not be affected.
- Do not use the device if you think it is damaged or if anything appears unusual.
- Do not store the unit under direct sunlight, at a high temperature, or in high humidity or dust. Performance may be degraded.
- Using the unit in the immediate vicinity of mobile phones, microwave appliances or other devices with strong electromagnetic fields may result in impaired functioning.
- Do not use this device close to strong electromagnetic fields, such as mobile telephones or radio installations. Keep a distance from such devices when using this unit.

## TEMPERATURE READINGS

Temperature readings vary from person to person. Although there is no one "normal" temperature reading, a temperature, taken orally, ranging between 37.4°C / 99.4°F and 38.8°C / 101.8°F is considered normal. A rectal temperature is generally 0.5°C / 1°F higher and an underarm (axillary) temperature will be 0.5°C / 1°F lower.

The best method to determine your own normal temperature is to use the thermometer when you are feeling well. Record your reading twice a day (early morning and late afternoon). Take the average of the two temperatures. This is considered your normal body temperature. Any variation from it may indicate illness and you should consult your physician.

## BEFORE USE

Your Vicks Fever InSight® Thermometer is a 3-in-1 thermometer for oral, rectal, or underarm use. Whichever method is used, it is recommended that for 15 minutes prior to use, you avoid eating or drinking, exercising, taking showers or baths or smoking.

## OPERATING YOUR VICKS FEVER INSIGHT® THERMOMETER

### 1. TURNING THE THERMOMETER ON

- Turn thermometer on by pressing the On/Off button (Fig. 1).
- A beep signal will sound and the display will flash Red/Yellow/Green for 1 second.
- Wait for the thermometer to get ready
  - The display will perform a self-test (Fig. 2).
  - The display will show last measurement for 2 seconds (Fig. 3).
  - After about 5 seconds, the display will show it's ready (Fig. 4).

### 2. SELECTING WHERE TO TAKE THE TEMPERATURE

#### Oral Use

- It is important to place the tip of the thermometer well under the tongue to get a good, accurate reading. (Fig. 5) Place sensor tip in areas marked with "✓".
- Keep mouth closed and sit still to help ensure an accurate measurement.
- When peak temperature is reached (usually within 12 seconds), 5 beep signals will sound and the "°C/°F" sign will stop flashing. The temperature measurement is now ready to read and will not change when the thermometer is removed.
- Wash the thermometer as recommended.

#### Under-the-arm (Axillary) Use

This method can be used for babies and young children.

- Wipe the underarm with a towel.
- Place the probe tip under the arm so the tip is touching the skin and position the arm next to the body (Fig. 6). With a young child, it may be helpful to hug the child to keep their arm next to their body. This ensures that the room air does not affect the reading.
- When peak temperature is reached (usually within 11 seconds but sometimes longer for this method), 5 beep signals will sound and the "°C/°F" sign will stop flashing. The temperature measurement is now ready to read and will not change when the thermometer is removed.

**Note: This method produces a temperature reading 0.5°C / 1°F lower than if the temperature were taken orally.**

- Wash the thermometer as recommended.

#### Rectal Use

Commonly used for babies, young children, or when it is difficult to take an oral or underarm temperature.

- Cover the thermometer's tip with a probe cover and lubricate with a water-soluble jelly for easier insertion. **DO NOT** use petroleum jelly. Read the label of the lubricant to be sure you are using the right kind.
- Lay the individual on his/her side. If the individual is an infant, the proper position for a baby is to lay on its stomach with legs hanging down, either across your knees or at the edge of a bed or changing table. This positions the infant's rectum for safe and easy insertion of the thermometer.
- Gently insert the tip of the thermometer **NO MORE THAN 1/2 inch** into the rectum. If you detect resistance, stop. Hold the thermometer in place during measurement.
- When the peak temperature is reached 5 beep signals will sound and the "°C" sign will stop flashing. The temperature is now ready to read and will not change when the thermometer is removed.

**Note: This method produces a temperature 1° higher than a temperature taken orally.**

- Dispose of used probe cover and wash the thermometer as recommended.

### 3. READING THE MEASUREMENT: FEVER INSIGHT® FEATURE

The Vicks Fever InSight® Thermometer comes with the patented "Fever InSight®" feature which offers a color indicator with the temperature reading. This feature will help you explain to your child the meaning of temperature. It also helps by distracting your child from possible fear associated with temperature taking, especially with young children.

- For readings between 35.8-37.4°C / 96.4-99.4°F, the screen will display green as the backlight color indicating the temperature is within the "normal" range.
- For readings between 37.5-38.5°C / 99.5-101.3°F, the screen will display yellow as the backlight color indicating a slight elevation in temperature.
- For readings 38.6°C / 101.4°F and above, the screen will display red as the backlight color indicating possible fever.

The Fever InSight® feature continues for about 5 seconds.

### 4. TURNING THE UNIT OFF

Turn the thermometer off by pressing the On/Off button. If you forget to turn it off, the thermometer will automatically turn off after approximately 1 1/2 minutes.

## CHANGING TEMPERATURE SCALE FROM °C TO °F OR °F TO °C

When the power is off, press and hold the On/Off button for 4 seconds, when the "M" symbol disappears, release the button and press the button to switch from °C to °F or °F to °C. After 3 seconds idle, the thermometer enters into the ready-for-measurement mode.

## CLEANING INSTRUCTIONS

- Clean the thermometer by washing it with soap and warm water or rubbing alcohol (70% isopropyl alcohol).
- Wipe dry with soft cloth.
- **DO NOT BOIL OR WASH IN DISHWASHER.** This will cause the thermometer to no longer function and will void the warranty.
- Do not wash with any thinner or chemical solvent.

## SPECIFICATION

Type:	Digital Thermometer (Predictive)
Battery Lifetime:	Approx. 2000 measurements.
Accuracy in water bath:	±0.1°C between 35.5°C / 42.0°C at room temperature of 21.6°C ±0.2°F between 96.0°F-107.6°F at room temperature of 71°F
Temperature Range:	32.0-43°C. If less than 32.0°C "Lo" is displayed. If more than 43°C "Hi" is displayed. 89.6°F-109.4°F. If less than 89.6°F "Lo" is displayed. If more than 109.4°F "Hi" is displayed.
Operating Temperature:	+10°C to +40°C (50°F-104°F), 15% to 95% RH
Storage Temperature:	-25°C to +60°C (-13°F-140°F), 15% to 95% RH
	If the device is not used within specified temperature and humidity ranges the technical accuracy of the measurement cannot be guaranteed.
Beeper:	The thermometer will beep for approximately 8 seconds when peak temperature is reached. For temperatures over 38.3°C/101.3°F, the audible Fever Alert will sound 5 times.
Error Messages:	The display will read "Err" when a malfunction occurs and the thermometer will power off.
Expected Service Life:	5 years

The stipulations of EU-Directive 93/42/EEC for Medical Devices Class IIa have been fulfilled.



Device standard: • EN 12470-3 • ASTM E1112 • ISO 80601-2-56 • IEC 60601-1 • IEC 60601-1-2(EMC) • IEC 60601-1-11  
• IP67: Dust tight, no ingress of dust; 2 to 8 hours. Protection against the effects of immersion in water between 15 cm and 1 m for 30 minutes.

## CHANGING THE BATTERIES

The Vicks Fever InSight® Thermometer features a long life battery. When the "V" appears and flashes in the center top of the display, the battery needs replacing.

1. Using a coin, unlock the battery cover and carefully remove the battery with a non-metallic instrument.
2. Replace the battery with a 3V, CR1632 lithium type or equivalent. Insert the new battery with the positive (+) side facing the battery cover.
3. Carefully put the battery cover back in place and, using a coin, tighten the cover.
4. Properly dispose of the old battery. Keep out of reach of young children, elderly and pets. Ensure the battery door is completely secure after replacing the battery.

**WARNING - KEEP BATTERIES OUT OF REACH OF CHILDREN.** Some parts are small enough to be swallowed. If you suspect a button battery has been ingested, call Poisons Information immediately on 13 11 26. Ensure battery compartment is correctly secured.

## LIFETIME WARRANTY

This product has been produced with the greatest care for normal, household use. Following directions carefully will ensure dependable operation. If for any reason (other than misuse or normal battery operation) you are dissatisfied with your Vicks Fever InSight® Thermometer, we will repair or replace it, at our option, at no charge during the lifetime warranty period. Modification to the product by the consumer is not authorised and voids the warranty.

If the thermometer does not function properly, first check the battery. Replace if necessary. Batteries and packaging are excluded from the warranty. If you experience unsatisfactory operation, please contact Felton Grimwade & Bosisto's. If repair is required under the warranty period, the product should be returned with proof of purchase to Felton Grimwade & Bosisto's service centre.

The purchaser assumes responsibility for the proper care and use of the product in accordance with the printed operating manual. The purchaser or user must make his or her judgment as to when to use the product and the length of use. Keep the operation manual available as it is your guide to safe, efficient operation.

Our goods come with guarantees that cannot be excluded under the Australian Consumer Law. You are entitled to a replacement or refund for a major failure and for compensation for any other reasonably foreseeable loss or damage. You are also entitled to have the goods repaired or replaced if the goods fail to be of acceptable quality and the failure does not amount to a major failure.

This product is manufactured by Microlife Intellectual Property GMBH for Helen of Troy Health and Home, under license from The Procter & Gamble Company.

Certain trademarks used under license from The Procter & Gamble Company or its affiliates. Fever InSight® is a registered trademark owned by Helen of Troy Limited and/or its affiliates.

© 2021 All rights reserved.

Batteries and electronic instruments must be disposed of in accordance with the locally applicable regulations, not with domestic waste.

**FGB** Distributed by: Felton Grimwade & Bosisto's Pty. Ltd.  
Natural Products 61-81 Clarinda Road (Po Box 74), Oakleigh South, Victoria 3167, Australia  
Free call: 1800 655 841 E-mail: cservice@fgb.com.au  
Website: www.fgb.com.au  
Australian sponsor: RQ Solutions Pty Ltd.  
Suite 1A Level 2, 802 Pacific Highway, Gordon NSW 2072  
Microlife Intellectual Property GMBH  
55 Northern Blvd., Suite 200 Great Neck, NY 11021  
Made and printed in China www.kaz.com/vicks

P/N: A005506RO DATE: 02NOV20



Important: Before first use, remove label from the display.



Fig. 1



Fig. 2



Fig. 3



Fig. 4



Fig. 5




Fig. 6

This page is only for legend box,  
Not need print this page

**Helen  
of Troy**

**Creative Department  
Artwork Specifications**

**Brand:** Vicks  
**Category:** Thermometer

**Model:** **V916C-AUSV1**

**Artwork Part #:** A005506R0

**Die Line Part #:** NA

**Oracle Desc. :** OWNER'S MANUAL, KAZ Artwork, V916C-AUS,  
ENGLISH ONLY, VELVET

**Subject:** OM

**Region:** AU

**Flat Size:** IN: **W4.75 x H14.5**  
MM: **W120.65 x H368.3**

**Scale:** 1/1

**Material:** 80g wood paper

**Page count:** NA

**Revision:** 2

**Date:** **02NOV20**

**Release Date:**

**Rerelease Date:**

**Colors:**

Dielines (Do not print)

Spot Colors

Cyan  
0%

Magenta  
0%

Yellow  
0%

Black  
100%

PMS Goe  
XXXX

PMS  
XXXX



Coating:  
Overall Gloss



Varnish:  
Spot UV

**Special Instructions:**

**Quality Requirement of Artwork and Quality  
Clarification Process of Artwork Printing:  
Meet Eng-QS-06&02**

Helen of Troy  
Creative Services  
Marlborough, MA 01752 USA  
+1 508 490 7000